


Serial No. C 116187

Weightron Bilanciai Ltd,
 Unit 4b, Block 4
 Westpark, Chelston
 Wellington TA21 9AD, UK
 Tel: 01823 662355
 Fax: 01823 662017
 www.weightron.com



Weightron Bilanciai Ltd,
 Titan Works
 Bridge Way off Broombank Road,
 Chesterfield S41 9QJ, UK
 Tel: 01246 260062
 Fax: 01246 260844
 www.weightron.com

CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

Kenny Waste
 Gibby Road
 Admanshaw
 M34 5HT
 LOCATION

Machine Details

Cust ref: B Doman
 Make: B D 700
 Model: B D 700
 Ser No: 120390078
 Capacity: 10000 kg /
 e/d: 20 kg /

Test Weights

| | | |
|---------------|-----------|-----------|
| Set No. | Cert No. | Date Cal |
| MS50 | 717211850 | APR 23 |
| Cert Interval | 12 Months | Next Cert |
| | May | 2025 |

| | | |
|----------|---------------|----------|
| 20 kg | 0-10000 | 20 kg |
| 40 kg | 10000-40000 | 40 kg |
| 10000 kg | 40000-100000 | 10000 kg |
| 20000 kg | 100000-200000 | 20000 kg |

Comments: Make up weight
 FLT-5640 kg
 MTO-19920

| | | |
|---|---|---|
| 1 | 3 | 5 |
| 2 | 4 | 6 |

AS FOUND LINEARITY & HYSTERESIS

| Load | Unloading | Dev | Loading | Dev |
|-------|-----------|-----|---------|-----|
| 0 | 0 | 0 | 0 | 0 |
| 4000 | 4000 | 0 | 4000 | 0 |
| 7000 | 7000 | 0 | 7000 | 0 |
| 10000 | 10000 | 0 | 10000 | 0 |
| 14000 | 14000 | 0 | 13980 | -20 |
| 16000 | 16000 | 0 | 16000 | 0 |
| 21640 | 21620 | -20 | 21620 | -20 |
| 19940 | 19920 | -20 | 19920 | -20 |
| 43580 | 43540 | -40 | 43540 | -40 |

DEFINITIVE LINEARITY & HYSTERESIS

| Load | Unloading | Dev | Loading | Dev |
|-------|-----------|-----|---------|-----|
| 0 | 0 | 0 | 0 | 0 |
| 4000 | 4000 | 0 | 4000 | 0 |
| 7000 | 7000 | 0 | 7000 | 0 |
| 10000 | 10000 | 0 | 10000 | 0 |
| 14000 | 14000 | 0 | 13980 | -20 |
| 16000 | 16000 | 0 | 16000 | 0 |
| 21640 | 21620 | -20 | 21620 | -20 |
| 19940 | 19920 | -20 | 19920 | -20 |
| 43580 | 43540 | -40 | 43540 | -40 |

Repeatability

| Load | Ind | Ind | Ind | Ind |
|-------|-------|-------|-------|-------|
| 43580 | 43580 | 43540 | 43540 | 43560 |

Repeatability

| Load | Ind | Ind | Ind | Ind |
|-------|-------|-------|-------|-------|
| 43580 | 43580 | 43540 | 43540 | 43560 |

Eccentric Load

| Reading | Dev | Reading | Dev |
|---------|-----|---------|-----|
| 9620 | -20 | 9620 | -20 |
| 9620 | -20 | 9640 | 0 |
| 9620 | -20 | 9640 | 0 |
| 9620 | -20 | 9640 | 0 |

Eccentric Load

| Reading | Dev | Reading | Dev |
|---------|-----|---------|-----|
| 9620 | -20 | 9620 | -20 |
| 9620 | -20 | 9640 | 0 |
| 9620 | -20 | 9640 | 0 |
| 9620 | -20 | 9640 | 0 |

Signature C1.3.88.1N

Name M Robert

Date 1/5/2024